

# Manual Supplement

Manual Title: 718 Users  
Print Date: July 1998  
Revision/Date: 4, 3/06

Supplement Issue: 7  
Issue Date: 2/24  
Page Count: 4

---

---

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 718  
CD Rev. & Date: 3, 2/2006  
CD PN: 1549626

## Change #1, 41517

On page 26, following the **Weight** specification add:

**Protection Class:** Pollution Degree II

## Change #2

On page 1, replace the fifth paragraph with the following:

Your Calibrator is supplied with a holster, two installed 9 V alkaline batteries, one set of TL75 test leads, one set of AC70A alligator clips, a Product Overview Manual, and a CD-Rom.

## Change #3, 59695

On page 1, under the **Pressure Calibrator** heading replace the first paragraph with:

The Calibrator makes 5-digit pressure readings in the following units: psi, inH<sub>2</sub>O at 4 °C, inH<sub>2</sub>O at 20 °C, kPa, cmH<sub>2</sub>O at 4 °C, cmH<sub>2</sub>O at 20 °C, bar, mbar, kg/cm<sup>2</sup>, inHg, mmHg, and inH<sub>2</sub>O at 60 °F.

On page 2, add the following to **Table 1**:

inH <sub>2</sub> O at 60 °F
-----------------------------


On page 25, replace the *Pressure Sensor Input* Table with:

*Pressure Sensor Input*

Model	Range	Max	Accuracy <sup>[1]</sup>	
			6 month	1 year
718 1G	-1 to 1 psi	5 psi	0.050 %	0.050 %
	-7 to 7 kPa	34.5 kPa		
718 30G	-12 to 30 psi	60 psi	0.025 %	0.035 %
	-83 to 207 kPa	413 kPa		
718 100G	-12 to 100 psi	200 psi	0.025 %	0.035 %
	-83 to 690 kPa	1.4 mPa		
718 300G	-12 to 300 psi	375 psi	0.035 %	0.050 %
	-83 to 2.1 mPa	2.6 mPa		
[1] In an RF field of 1 V/m to 3 V/m between the frequencies of 1.4 GHz to 2.0 GHz, add 120 counts to specified accuracy.				

## Change #4, 59845

In the **Specification** section, add the following:

Specifications apply to the 718 Pressure Calibrators with firmware version 4.0 or greater. To verify the firmware version, press and hold  and power on the Calibrator.

## Change #5, 67387

On page 26, under **General Specifications**, add:

Electromagnetic Compatibility .....Applies to use in Korea only. Class A Equipment (Industrial Broadcasting & Communication Equipment)

This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.

## Change #6, 37

On page 1, replace the fifth paragraph with:

The Calibrator is supplied with a holster, two installed 9 V alkaline batteries, one set of TL75 test leads, one set of AC70A alligator clips, a *Product Overview Manual*, and a CD-Rom.

## Change #7, Asana 261

On page 22, Table 8, change these items:

Replace the AC72 row with:

<b>AC175</b>	Alligator Clip, Black	4239092	1
	Alligator Clip, Red	4239050	1

On the MP6 row:

2558508 only applies to 30G and 100G

300G should be part number 2632057

On page 24, Figure 9:

Replace AC72 with AC175.