

# Manual Supplement

Manual Title:	772/773 Calibration	Supplement Issue:	1
Part Number:	Web Only	Issue Date:	1/22
Print Date:	October 2009	Page Count:	1
Revision/Date:	1, 5/19		

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

**Change #1, 241**

On page 12, replace this table:

***mA DC Clamp Measure Accuracy Tests***

Step	Calibrator Output	UUT Meter Reading Limit	
		Low	High
1	4.00 mA	3.94 mA	4.06 mA
2	-4.00 mA	-4.06 mA	-3.94 mA
3	12.00 mA	11.93 mA	12.07 mA
4	-12.00 mA	-12.07 mA	-11.93 mA
5	20.00 mA	19.91 mA	20.09 mA
6	-20.00 mA	-20.09 mA	-19.91 mA
7	100.0 mA	98.5 mA	101.5 mA
8	-100.0 mA	-101.5 mA	-98.5 mA