

Manual Supplement

Manual Title: 789/787B Users
Print Date: August 2002
Revision/Date: 4, 1/17

Supplement Issue: **3**
Issue Date: 10/21
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 789 ProcessMeter™
CD Rev. & Date: 3, 3/13
CD PN: 1636493

Change #1, 173

On page 40, under **DC Current Measurement**, replace the Note with:

Note: 440 mA continuous, 1 A 30 seconds maximum on, 5 minutes off

Change #2, 633

On page 40, under **AC Current Measurement**, replace the **Range** with:

Range

45 Hz to 500 Hz

Change #3, 210

On page 36, in Table 13, item 9, change the **789 PN**:

from:

1622855

to:

1622862