

FLUKE®

718Ex 30G/100G/300G

Pressure Calibrator

Users Manual

May 2004 Rev. 4, 1/25

© 2004-2025 Fluke Corporation. All rights reserved. Specifications are subject to change without notice.

All product names are trademarks of their respective companies.

LIMITED WARRANTY AND LIMITATION OF LIABILITY

Each Fluke product is warranted to be free from defects in material and workmanship under normal use and service. The warranty period is three years (one year for pump assembly) and begins on the date of shipment. Parts, product repairs, and services are warranted for 90 days. This warranty extends only to the original buyer or end-user customer of a Fluke authorized reseller, and does not apply to fuses, disposable batteries, or to any product which, in Fluke's opinion, has been misused, altered, neglected, contaminated, or damaged by accident or abnormal conditions of operation or handling. Fluke warrants that software will operate substantially in accordance with its functional specifications for 90 days and that it has been properly recorded on non-defective media. Fluke does not warrant that software will be error free or operate without interruption.

Fluke authorized resellers shall extend this warranty on new and unused products to end-user customers only but have no authority to extend a greater or different warranty on behalf of Fluke. Warranty support is available only if product is purchased through a Fluke authorized sales outlet or Buyer has paid the applicable international price. Fluke reserves the right to invoice Buyer for importation costs of repair/replacement parts when product purchased in one country is submitted for repair in another country.

Fluke's warranty obligation is limited, at Fluke's option, to refund of the purchase price, free of charge repair, or replacement of a defective product which is returned to a Fluke authorized service center within the warranty period.

To obtain warranty service, contact your nearest Fluke authorized service center to obtain return authorization information, then send the product to that service center, with a description of the difficulty, postage and insurance prepaid (FOB Destination). Fluke assumes no risk for damage in transit. Following warranty repair, the product will be returned to Buyer, transportation prepaid (FOB Destination). If Fluke determines that failure was caused by neglect, misuse, contamination, alteration, accident, or abnormal condition of operation or handling, including overvoltage failures caused by use outside the product's specified rating, or normal wear and tear of mechanical components, Fluke will provide an estimate of repair costs and obtain authorization before commencing the work. Following repair, the product will be returned to the Buyer transportation prepaid and the Buyer will be billed for the repair and return transportation charges (FOB Shipping Point).

THIS WARRANTY IS BUYER'S SOLE AND EXCLUSIVE REMEDY AND IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. FLUKE SHALL NOT BE LIABLE FOR ANY SPECIAL, INDIRECT, INCIDENTAL, OR CONSEQUENTIAL DAMAGES OR LOSSES, INCLUDING LOSS OF DATA, ARISING FROM ANY CAUSE OR THEORY.

Since some countries or states do not allow limitation of the term of an implied warranty, or exclusion or limitation of incidental or consequential damages, the limitations and exclusions of this warranty may not apply to every buyer. If any provision of this Warranty is held invalid or unenforceable by a court or other decision-maker of competent jurisdiction, such holding will not affect the validity or enforceability of any other provision.

11/99

Fluke Corporation
6920 Seaway Blvd.
Everett, WA 98203
U.S.A.

Fluke Europe B.V.
P.O. Box 1186
5602 BD Eindhoven
The Netherlands

Table of Contents

| Title | Page |
|--|------|
| Introduction | 1 |
| Contact Fluke | 2 |
| Safety Information | 2 |
| Getting Acquainted with the Calibrator | 3 |
| Power Saver | 6 |
| Zeroing with Absolute Pressure Modules | 6 |
| Calibrating a P/I Transmitter | 8 |
| Using the Internal Pump..... | 8 |
| Using an External Pump | 12 |
| External Fluke Pressure Module Compatibility | 14 |
| Switch Test..... | 15 |
| Maintenance..... | 15 |
| In Case of Difficulty..... | 16 |
| Cleaning..... | 16 |
| Cleaning the Pump Valve Assembly..... | 16 |
| Product Disposal..... | 17 |
| Calibration..... | 17 |

| | |
|--|----|
| Replacing the Battery | 17 |
| Parts and Accessories | 19 |
| Specifications | 20 |
| Pressure Sensor Input | 20 |
| Pressure Sensor Range and Resolution | 21 |
| Pressure Module Input | 21 |
| DC mA Input | 21 |
| General Specifications | 22 |

718Ex 30G/100G/300G Pressure Calibrator

Introduction

Warning

Read Safety Information before using the Calibrator.

The Fluke Model 718Ex 30G, 718Ex 100G, and 718Ex 300G Pressure Calibrators (hereafter called Calibrator) can do the following:

- Calibrate P/I (pressure to current) transmitters.
- Measure pressure via a 1/8-inch NPT pressure fitting and an internal pressure sensor or via Fluke 750PEX Series Pressure Modules.
- Measure current up to 24 mA.
- Simultaneously display pressure and current measurements.
- Perform switch testing.

The Calibrator is for use ONLY in Ex-hazardous areas.

The Calibrator makes 5-digit pressure readings in the following units: psi, inH₂O at 4 °C, inH₂O at 20 °C, kPa, cmH₂O at 4 °C, cmH₂O at 20 °C, bar, mbar, kg/cm², inHg, and mmHg. Full-scale pressure sensor input is as follows:

- Model 718Ex 30G: 30 psi (206.85 kPa, 2.0685 bar). **OL** appears at 33 psi.
- Model 718Ex 100G: 100 psi (689.5 kPa, 6.895 bar). **OL** appears at 120 psi.
- Model 718EX 300G: 300 psi (2068 kPa, 20.68 bar). **OL** appears at 360 psi.

The Calibrator measures pressure sensor inputs in the units shown under Pressure Sensor Range and Resolution.

For Pressure Modules, full-scale readings for all pressure ranges can be made in psi, kPa, and inHg units. To avoid

display overflow, full-scale readings are limited to 1000 psi in cmH₂O, mbar, and mmHg units, and 3000 psi in inH₂O units. Pressures of at least 15 psi must be measured for meaningful readings in bar and kg/cm² units.

The Calibrator is supplied with:

- a holster
- one installed 9 V battery
- one set of TL75 test leads
- one set of AC175 alligator clips
- a Control Drawing

If the Calibrator is damaged or something is missing, contact the place of purchase immediately. Contact a Fluke distributor for information about accessories. See *Contact Fluke*. To order replacement parts or spares, see *Parts and Accessories*.

Contact Fluke

Fluke Corporation operates worldwide. For local contact information, go to our website:

www.fluke.com.

To register your product, or to view, print, or download the latest manual or manual supplement, go to our website.

+1-425-446-5500

<mailto:fluke-info@fluke.com>

Safety Information


General Safety Information is in the printed Safety Information document that ships with the Product and at www.fluke.com. More specific safety information is listed where applicable.

A **Warning** identifies conditions and actions that pose hazard(s) to the user; a **Caution** identifies conditions and actions that may damage the Calibrator or the equipment under test.

⚠ Caution

To avoid mechanically damaging the Calibrator:

- **Do not apply torque between the pressure fitting and the Calibrator case. See Figure 1 for the proper use of tools.**
- **To avoid damage to the pump, use with dry air and non-corrosive gases only.**



wh001f.emf

Figure 1. Connection Technique

Getting Acquainted with the Calibrator












Press **Ⓢ** to turn the Calibrator on and off. The Calibrator displays pressure and current measurements simultaneously. See Table 1.

The upper part of the display shows the applied pressure or vacuum. Vacuum is shown as a negative value. Press **UNITS** to select a different unit. When cycling the power off and on, the Calibrator retains the unit last used.

The lower part of the display shows the current (up to 24 mA) applied to the current (mA) inputs.

Pushbutton operation is described in Table 2. Pump features are shown in Figure 2 described in Table 1.

Table 2. Pushbutton Functions




| Pushbutton | Description |
|---|---|
|  | Press to select a different pressure unit. All units are available when the pressure sensor input is used. For higher pressure module inputs, inappropriate (out-of-range) units are not available. |
|  | Turns pressure reading damping on and off. With damping on, the Calibrator averages several measurements before displaying a reading. |
|  | Press to zero the pressure display. Vent pressure to atmosphere before pressing this pushbutton. For an Absolute Pressure Module, see <i>Zeroing with Absolute Pressure Modules</i> . |
|  | Press to read the minimum pressure and current readings since power was turned on or  was pressed. Press again to read the maximum pressure and current readings since power was turned on or  was pressed. |
|  | Use for pressure switch test. See <i>Switch Test</i> . |
|  | Press to clear the MIN, MAX, and switch test memories. |
|  | Press  to freeze the display. The HOLD symbol appears on the display. Press  again to resume normal operation. |

Power Saver

The Calibrator automatically turns off after 30 minutes of inactivity. To reduce this time or disable this feature:

1. With the Calibrator OFF, press .






P.S. xx is displayed, where **xx** is the turn-off time in minutes. **OFF** means the power saver is disabled.

2. Press  to decrease or  to increase the turn-off time.
3. To disable, press  until the display shows **OFF**.

The Calibrator resumes normal operation after 2 seconds.

Zeroing with Absolute Pressure Modules

For zeroing, adjust the Calibrator to read a known pressure. This can be barometric pressure, if it is accurately known. An accurate pressure standard can also apply a pressure within range for any Absolute Pressure Module. Adjust the Calibrator reading as follows:

1. Press and hold .
2. Press  to increase or  to decrease the Calibrator reading to equal the applied pressure.
3. Release  to exit the zeroing procedure.
4. Press the  button to convert to any convenient measurement display unit.




Figure 3. Pump Features

alz009f.emf

Table 3. Pump Features

| Item | Description |
|------|---|
| ① | Pressure Vacuum Switch - Rotate forward (clockwise) for pressure, backward (counter-clockwise) for vacuum. |
| ② | Pressure Vacuum Release Control - Rotate fully backward (counter-clockwise) to release all pressure or vacuum. (Rotate slightly for partial release.) Rotate fully forward (clockwise) to close valve. |
| ③ | Fine Adjustment Knob - Rotate either direction for precise adjustment of applied pressure or vacuum. Full rotation is about 30 turns. |
| ④ | Internal Pump - Increase pressure on the inward stroke. In vacuum mode, decrease pressure on the outward stroke. |

Calibrating a P/I Transmitter

To calibrate a P/I (pressure to current) transmitter, apply a pressure to the transmitter and measure the transmitter's current loop output. Pressure can be applied with the Calibrator's internal pump or with an external pump.

⚠ Warning

To avoid a violent release of pressure or vacuum, always depressurize the system slowly using the pressure/vacuum release control before detaching any pressure line.

When measuring the pressure of potentially hazardous gases, care must be taken to minimize the possibility of leakage:

- **Confirm that all pressure connections are properly sealed.**
- **Confirm that the Pressure/Vacuum Release Control is in the closed position (fully clockwise) and the Pressure/Vacuum switch is in the + position (fully clockwise).**

- **If the Calibrator has been dropped or subjected to rough handling, remove the Calibrator to a safe area and check for leaks to confirm the integrity of the internal pneumatic components.**

Using the Internal Pump

The internal pump can provide 30 psi (2.0685 bar) for Model 718Ex 30G, 100 psi (6.895 bar) for Model 718Ex 100G, or 300 psi (20.68 bar) for Model 718Ex 300G.

The preferred use for the internal pump is shown in Figure 4, where the Calibrator displays pressure measured with the internal sensor and provided by the internal pump.

The internal pump can also be used with certain Fluke 750PEX Series Pressure Modules. In this case, pressure measured by the pressure module is displayed by the Calibrator. Appropriate pressure modules for each Calibrator model are identified in Table 4. Figure 4 shows the internal pump being used with a pressure module.

Note

To prevent misleading readings when using the Calibrator's internal pressure sensor, do not connect a pressure module at the Calibrator. To avoid misleading readings, disconnect the pressure module connector at the Calibrator.

⚠ Warning

If both a pressure module and the internal sensor are connected, the Calibrator displays ONLY the pressure module measurement.

To use the Calibrator's internal pump, refer to Figure 4 and perform the following steps:

1. Depressurize the line before connecting the Calibrator.
2. Connect the pressure transmitter under test to the Calibrator internal sensor as shown in Figure 4 (for internal pressure sensor measurements) or Figure 5 (for pressure module measurements.)

Note


To avoid leaks, use PTFE tape or similar sealant on all pressure connections.

3. Make sure the pressure/vacuum switch on the Calibrator is in the desired position. Forward (clockwise) is for pressure; backward (counter-clockwise) is for vacuum.
4. Turn the pressure/vacuum release control backward (counter-clockwise) to vent pressure/vacuum from the pump.
5. Press **ZERO** to zero the pressure display.
6. Turn the fine adjustment knob to mid-range.
7. Turn the pressure/vacuum release control forward (clockwise) to close the release valve.
8. Work the pump handle in and out to apply incrementally larger pressure/vacuum changes. Shorten the stroke to apply smaller increments of pressure/vacuum change.
9. To make very small pressure/vacuum changes, use the fine adjustment knob.

Note

This knob adjusts a small internal reservoir to vary the total volume. With larger external pressure/vacuum volumes, this control will adjust pressure or vacuum within a smaller range.

10. Depressurize the system before disconnecting the pressure line.



alz002f.emf

Figure 4. Internal Pressure Sensor with Internal Pump




Figure 5. Pressure Module with Internal Pump

alz010f.emf

Table 4. Recommended Pressure Modules

| Pressure Module | External Pump | Internal Pump | | |
|-----------------|---------------------|---------------|------------|------------|
| | 718Ex 30G/100G/300G | 718Ex 30G | 718Ex 100G | 718Ex 300G |
| 750P01Ex | X | X | X | X |
| 750P24Ex | X | X | X | X |
| 750P05Ex | X | X | X | X |
| 750P06Ex | X | - | X | X |
| 750P27Ex | X | - | - | X |
| 750P09Ex | X | - | - | - |
| 750PA4Ex | X | X | X | X |
| 750P29Ex | X | - | - | - |

Using an External Pump

Warning

To avoid damage to the Calibrator and possible release of pressure, do not connect the internal sensor to an external pressure source that exceeds 30 psi for Model 718Ex 30G, 100 psi for Model 718Ex 100G, or 300 psi for Model 718Ex 300G.

To develop higher pressure or vacuum, use an external pump. Use a Fluke 750PEX Pressure Module connected to the pressure module input on the Calibrator. Pressure modules are listed in Table 4. Make overall connections as shown in Figure 6.

Refer to setup and operating instructions included with the pressure module and pump.




Figure 6. Pressure Module with External Pump

alz006f.emf

External Fluke Pressure Module Compatibility

If inappropriate units are selected, the output of Fluke 750PEx pressure modules can cause the Calibrator display to overflow (OL), or display values that are too low to be read. Refer to Table 5 for appropriate unit and range compatibility.

Table 5. Fluke Pressure Module Compatibility


| Pressure Unit | Module Compatibility |
|--------------------|----------------------------------|
| psi | Available on all pressure ranges |
| inH ₂ O | All ranges through 3000 psi |
| cmH ₂ O | All ranges through 1000 psi |
| bar | 15 psi and above |
| mbar | All ranges through 1000 psi |
| kPa | Available on all pressure ranges |
| inHg | Available on all pressure ranges |
| mmHg | All ranges through 1000 psi |
| kg/cm ² | 15 psi and above |

Switch Test

To perform a switch test, follow these steps:




Note

This example used a normally closed switch. The procedure is the same for an open switch but the display reads OPEN instead of CLOSE.

1. Connect the Calibrator mA and COM terminals to the switch using the pressure switch terminals and connect the pump from the Calibrator to the pressure switch. The polarity of the terminals does not matter.
2. Make sure the vent on the pump is open and zero the Calibrator if necessary. Close the vent after zeroing the Calibrator.
3. Press  to enter pressure switch test mode. The Calibrator will display CLOSE instead of a mA measurement.
4. Apply pressure with the pump slowly until the switch opens.

Note

In the switch test mode, the display update rate is increased to help capture changing pressure inputs. Even with this enhanced sample rate, pressuring the device under test should be done slowly to ensure accurate readings.

5. OPEN is displayed once the switch is open. Bleed the pump slowly until the pressure switch closes. RCL appears on the display.
6. Press  to read the pressure values for when the switch opened, for when it closed, and for the deadband.
7. Hold  for three seconds to exit the switch test or press  to reset the switch test.

Maintenance

Warning

To avoid personal injury, or sudden release of pressure, review Safety Information before proceeding.

For maintenance procedures not described in this manual, or if the Calibrator needs repair, contact a Fluke Service Center. See *Contact Fluke*.

In Case of Difficulty

- After removing the Calibrator from the Ex-hazardous area, check the battery, test leads, pressure module, and pressure tubing. Follow replacement and connection instructions properly.
- Review this manual and control drawing to make sure the Calibrator is used correctly.

If the Calibrator needs repair, and the Calibrator is under warranty, see the warranty statement for terms. If the warranty has lapsed, the Calibrator can be repaired and returned for a fixed fee.

Cleaning

Periodically wipe the case with a damp cloth; do not use abrasives or solvents.

Cleaning the Pump Valve Assembly

1. Using a small screwdriver, remove the two valve retention caps located in the oval-shaped opening on the back side of the Calibrator.
2. Gently remove the spring and o-ring assembly.
3. Set aside the valve assemblies in a safe area and clean out the valve body using a cotton swab soaked in IPA (isopropyl alcohol).

4. Repeat this process several times using a new cotton swab each time until there is no remaining residue.
5. Pump the unit several times and check again for residue.
6. Clean the o-ring assembly and the o-ring on the retention caps with IPA and inspect the o-rings closely for any cuts, nicks, or wear. Replace if needed.
7. Inspect the springs for wear or loss of tension. They should be approximately 8.6 mm long in the relaxed state. If they are shorter than that, they may not allow the o-ring to seat properly. Replace if needed.
8. Once all parts have been cleaned and inspected, reinstall the o-ring and spring assemblies into the valve body.
9. Reinstall the retention caps and gently tighten the cap.
10. Seal the output of the Calibrator and pump up the unit to at least 50 % its rated pressure.
11. Release the pressure and repeat several times to ensure that the o-rings seat properly.

The Calibrator is now ready to use.

Product Disposal

Dispose of the Product in a professional and environmentally sound manner:


- Delete personal data on the Product before disposal.
- Remove batteries that are not integrated into the electrical system before disposal and dispose of batteries separately.
- If this Product has an integral battery, put the entire Product in the electrical waste.

Calibration


Fluke recommends that the Calibrator be calibrated once yearly to ensure that it performs according to its specifications. A calibration manual is available. See *Contact Fluke*.


Replacing the Battery

⚠ Warning

- **To avoid false readings, which could lead to personal injury, replace the battery as soon as the battery indicator  appears.**
- **Remove the Calibrator from the Ex-hazardous area before opening the battery door.**

- **Use only the battery types listed in the Approved Battery Table in the *Safety Information*.**

When  appears on the display, replace the 9 V battery. Refer to Figure 7.



alz008f.emf

Figure 7. Battery Replacement

Parts and Accessories

Refer to Table 6 for a list of replacement parts and accessories.

Table 6. Replacement Parts and Accessories

| Model No. | Description | Part | Qty |
|-----------|------------------------------------|---|-----|
| AC175 | Alligator clip, Black | 4239092 | 1 |
| | Alligator clip, Red | 4239050 | 1 |
| BT1 | 9 V battery, Carbon Zinc, IEC 6F22 | 4982400 or see Battery Table in the Safety Information | 1 |
| Holster | Holster, Red | 2096118 | 1 |
| - | Battery Door Assembly | 2117013 | 1 |
| TL75 | Test Lead Set | 855742 | 1 |
| | Cap, Black | 3986568 | 1 |
| | Cap, Red | 3995524 | 1 |
| - | 71X Series Calibration Manual | See www.fluke.com | Opt |
| - | 718Ex Control Drawing | 2117024 | 1 |


Specifications

Specifications are based on a one year calibration cycle and apply for ambient temperature from +18 °C to +28 °C unless stated otherwise. Counts are the number of increments or decrements of the least significant digit.

Pressure Sensor Input

| Model | Range | Accuracy | Max Non-destructive Pressure |
|--|---|--------------------------------|------------------------------|
| 30G | -12 psi to 30 psi (-83 kPa to 207 kPa) | ±0.025 % of range (6 month) | 60 psi (413 kPa) |
| 100G | -12 psi to 100 psi (-83 kPa to 690 kPa) | ±0.035 % of range (1 year) | 200 psi (1.4 MPa) |
| 300G | -12 psi to 300 psi (-83 kPa to 2068 kPa) | ±0.05 % of range (1 year) | 375 psi (2.6 MPa) |
| <p><i>Temperature coefficient: 0.01 % of range per °C for temperature ranges -10 °C to 18 °C and 28 °C to 55 °C.</i></p> <p><i>Where in a 3 V/m radiated EM field ≥350 MHz, Pressure Accuracy is 1 % of range.</i></p> | | | |

Note

Specifications apply to the Product with firmware version 2.0 or greater. To verify the firmware version, push and hold  and power on the Calibrator.

Pressure Sensor Range and Resolution

| Displayed Pressure Units | Model 718Ex 30G Range and Resolution | Model 718Ex 100G Range and Resolution | Model 718Ex 300G Range and Resolution |
|--------------------------------|--|--|--|
| psi | -12.000 to 30.000 psi | -12.00 to 100.00 psi | -12.00 to 300.00 psi |
| inH ₂ O at 4 °C | -332.16 to 830.40 inH ₂ O | -332.2 to 2768.0 inH ₂ O | -332.2 to 8304 inH ₂ O |
| inH ₂ O at 20 °C | -332.75 to 831.87 inH ₂ O | -332.8 to 2772.9 inH ₂ O | -332.8 to 8318.7 inH ₂ O |
| cmH ₂ O at 4 °C | -843.6 to 2109.0 cmH ₂ O | -843.6 to 7030.0 cmH ₂ O | -843.6 to 21090 cmH ₂ O |
| cmH ₂ O at 20 °C | -845.2 to 2113.0 cmH ₂ O | -845.2 to 7043.0 cmH ₂ O | -845.2 to 21129 cmH ₂ O |
| bar | -0.8274 to 2.0685 bar | -0.8274 to 6.8950 bar | -0.8274 to 20.685 bar |
| mbar | -827.4 to 2068.5 mbar | -827.4 to 6895.0 mbar | -827.4 to 20685 mbar |
| kPa | -82.74 to 206.85 kPa | -82.74 to 689.50 kPa | -82.74 to 2068.5 kPa |
| inHg | -24.432 to 61.080 inHg | -24.43 to 203.60 inHg | -24.43 to 610.8 inHg |
| mmHg | -620.6 to 1551.4 mmHg | -620.6 to 5171.5 mmHg | -620.6 to 15514.5 mmHg |
| kg/cm ² | -0.8437 to 2.1090 kg/cm ² | -0.8437 to 7.0306 kg/cm ² | -0.8437 to 21.0918 kg/cm ² |

Pressure Module Input

| Range | Resolution | Accuracy |
|---------------------------------|------------|----------|
| (determined by Pressure Module) | | |

DC mA Input

| Range | Resolution | Accuracy, ±(% of Reading + Counts) |
|--|------------|------------------------------------|
| 24 mA | 0.001 mA | 0.02 + 2 |
| <p><i>Temperature coefficient: 0.005 % of range per °C for temperature ranges -10 °C to 18 °C and 28 °C to 55 °C.</i></p> <p><i>Where in a 3 V/m radiated EM field ≥350 MHz, add 0.1 % to mA input accuracy.</i></p> | | |

General Specifications

Safety Specifications are in the *Safety Information* that ships with the Product.

Entity Parameters mA Jack Input:

| Vi, Ui | Ii | Pi | Ci | Li |
|--------|-------|--------|-----------|------|
| 30 V | 80 mA | 0.60 W | 0 μ F | 0 mH |

Entity Parameters mA Jack Output:

| Vo, Uo | Io | Po | Co | | | Lo | | |
|--------|--------|--------|--------------|-------------|--------------|--------|--------|----------|
| | | | IIC | IIB | IIA | IIC | IIB | IIA |
| 7.14 V | 1.2 mA | 2.0 mW | 13.5 μ F | 240 μ F | 1000 μ F | 24.7 H | 98.7 H | 197.54 H |

Pressure Module Output Circuit:

| Vo, Uo | Io | Po | Co | | | Lo | | |
|--------|--------|--------|--------------|-------------|--------------|---------|---------|----------|
| | | | IIC | IIB | IIA | IIC | IIB | IIA |
| 7.14 V | 123 mA | 218 mW | 13.5 μ F | 240 μ F | 1000 μ F | 2.38 mH | 9.54 mH | 19.08 mH |

Power requirements: See Approved Batteries in the *Safety Information*.

Size: 66 mm H x 94 mm W x 216 mm L (2.60 in H x 3.70 in W x 8.5 in L)

Weight: 992 g (35 oz)